

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
 Data File : VX025381.D
 Acq On : 29 Nov 2021 19:42
 Operator : JC/MD
 Sample : M4837-20 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20937

Quant Time: Nov 30 04:04:43 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	132253	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	216623	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	201611	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94236	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72509	52.014	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 104.020%			
35) Dibromofluoromethane	5.391	113	66303	49.885	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 99.760%			
50) Toluene-d8	8.653	98	248403	49.707	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 99.420%			
62) 4-Bromofluorobenzene	11.085	95	90297	47.762	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 95.520%			
Target Compounds						Qvalue
13) Acrolein	2.245	56	1190	18.633	ug/l #	78
16) Acetone	2.380	43	32793	42.335	ug/l	91
25) 2-Butanone	4.568	43	3054	2.727	ug/l	99
64) Tetrachloroethene	9.275	164	883	0.531	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
Data File : VX025381.D
Acq On : 29 Nov 2021 19:42
Operator : JC/MD
Sample : M4837-20 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20937

Quant Time: Nov 30 04:04:43 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

