

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
 Data File : VX025383.D
 Acq On : 29 Nov 2021 20:28
 Operator : JC/MD
 Sample : M4837-08 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20928

Quant Time: Nov 30 04:05:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126414	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206040	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	190211	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90648	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70553	52.948	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.900%	
35) Dibromofluoromethane	5.391	113	62607	49.523	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.040%	
50) Toluene-d8	8.647	98	240423	50.581	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 101.160%	
62) 4-Bromofluorobenzene	11.079	95	86206	47.941	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 95.880%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	2976	4.019	ug/l #	85
40) Benzene	6.043	78	7560	1.388	ug/l #	75
52) Toluene	8.720	92	5584	1.571	ug/l	92
68) m/p-Xylenes	10.305	106	1432	0.523	ug/l	94
69) o-Xylene	10.646	106	714	0.267	ug/l	95
70) Styrene	10.659	104	2984	0.675	ug/l #	89
95) Naphthalene	13.780	128	5088	0.460	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
Data File : VX025383.D
Acq On : 29 Nov 2021 20:28
Operator : JC/MD
Sample : M4837-08 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20928

Quant Time: Nov 30 04:05:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

