

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
 Data File : VX025384.D
 Acq On : 29 Nov 2021 20:52
 Operator : JC/MD
 Sample : M4837-10 20X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 20930

Quant Time: Nov 30 04:05:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	125468	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206187	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	183242	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	84682	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69607	52.632	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 105.260%		
35) Dibromofluoromethane	5.391	113	63079	49.861	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.720%		
50) Toluene-d8	8.653	98	236491	49.719	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 99.440%		
62) 4-Bromofluorobenzene	11.085	95	82965	46.105	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 92.220%		
Target Compounds						Qvalue
16) Acetone	2.380	43	2127	2.894	ug/l	95
40) Benzene	6.037	78	1350	0.248	ug/l #	11
52) Toluene	8.726	92	1065	0.299	ug/l	98
70) Styrene	10.659	104	2125	0.499	ug/l	95
95) Naphthalene	13.780	128	3336	0.210	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
Data File : VX025384.D
Acq On : 29 Nov 2021 20:52
Operator : JC/MD
Sample : M4837-10 20X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
20930

Quant Time: Nov 30 04:05:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

