

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX112921\
 Data File : VX025385.D
 Acq On : 29 Nov 2021 21:15
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 30 04:05:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	50.000	50.000	0.0	76	0.00
2 T	Dichlorodifluoromethane	50.000	56.884	-13.8	76	0.00
3 P	Chloromethane	50.000	47.731	4.5	70	0.00
4 C	Vinyl Chloride	50.000	49.039	1.9#	71	0.00
5 T	Bromomethane	50.000	50.934	-1.9	76	0.00
6 T	Chloroethane	50.000	46.086	7.8	73	0.00
7 T	Trichlorofluoromethane	50.000	50.395	-0.8	74	0.00
8 T	Diethyl Ether	50.000	49.937	0.1	76	0.00
9 T	1,1,2-Trichlorotrifluoroeth	50.000	53.081	-6.2	81	0.00
10 T	Methyl Iodide	50.000	57.798	-15.6	80	0.00
11 T	Tert butyl alcohol	250.000	221.480	11.4	65	0.00
12 CM	1,1-Dichloroethene	50.000	52.044	-4.1#	78	0.00
13 T	Acrolein	250.000	355.990	-42.4#	104	0.00
14 T	Allyl chloride	50.000	47.915	4.2	76	0.00
15 T	Acrylonitrile	250.000	254.621	-1.8	76	0.00
16 T	Acetone	250.000	204.877	18.0	62	0.00
17 T	Carbon Disulfide	50.000	43.995	12.0	69	0.00
18 T	Methyl Acetate	50.000	50.509	-1.0	79	0.00
19 T	Methyl tert-butyl Ether	50.000	52.576	-5.2	81	0.00
20 T	Methylene Chloride	50.000	49.730	0.5	81	0.00
21 T	trans-1,2-Dichloroethene	50.000	51.066	-2.1	79	0.00
22 T	Diisopropyl ether	50.000	52.208	-4.4	81	0.00
23 T	Vinyl Acetate	250.000	259.803	-3.9	79	0.00
24 P	1,1-Dichloroethane	50.000	52.704	-5.4	81	0.00
25 T	2-Butanone	250.000	237.553	5.0	71	0.00
26 T	2,2-Dichloropropane	50.000	48.474	3.1	74	0.00
27 T	cis-1,2-Dichloroethene	50.000	52.546	-5.1	80	0.00
28 T	Bromochloromethane	50.000	52.280	-4.6	79	0.00
29 T	Tetrahydrofuran	250.000	247.321	1.1	76	0.00
30 C	Chloroform	50.000	53.453	-6.9#	82	0.00
31 T	Cyclohexane	50.000	49.867	0.3	77	0.00
32 T	1,1,1-Trichloroethane	50.000	52.969	-5.9	81	0.00
33 S	1,2-Dichloroethane-d4	50.000	53.591	-7.2	83	0.00
34 I	1,4-Difluorobenzene	50.000	50.000	0.0	77	0.00
35 S	Dibromofluoromethane	50.000	52.654	-5.3	82	0.00
36 T	1,1-Dichloropropene	50.000	49.867	0.3	78	0.00
37 T	Ethyl Acetate	50.000	49.782	0.4	78	0.00
38 T	Carbon Tetrachloride	50.000	51.620	-3.2	80	0.00
39 T	Methylcyclohexane	50.000	48.883	2.2	75	0.00
40 TM	Benzene	50.000	51.000	-2.0	80	0.00
41 T	Methacrylonitrile	50.000	49.232	1.5	80	0.00
42 TM	1,2-Dichloroethane	50.000	53.411	-6.8	83	0.00
43 T	Isopropyl Acetate	50.000	49.778	0.4	78	0.00
44 TM	Trichloroethene	50.000	51.108	-2.2	79	0.00
45 C	1,2-Dichloropropane	50.000	51.585	-3.2#	79	0.00
46 T	Dibromomethane	50.000	52.341	-4.7	81	0.00
47 T	Bromodichloromethane	50.000	50.794	-1.6	79	0.00
48 T	Methyl methacrylate	50.000	48.807	2.4	77	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
 Data File : VX025385.D
 Acq On : 29 Nov 2021 21:15
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 30 04:05:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	1000.000	922.013	7.8	69	0.00
50 S	Toluene-d8	50.000	51.518	-3.0	79	0.00
51 T	4-Methyl-2-Pentanone	250.000	248.503	0.6	77	0.00
52 CM	Toluene	50.000	51.143	-2.3#	79	0.00
53 T	t-1,3-Dichloropropene	50.000	50.426	-0.9	78	0.00
54 T	cis-1,3-Dichloropropene	50.000	49.966	0.1	77	0.00
55 T	1,1,2-Trichloroethane	50.000	53.639	-7.3	83	0.00
56 T	Ethyl methacrylate	50.000	51.340	-2.7	79	0.00
57 T	1,3-Dichloropropane	50.000	52.129	-4.3	81	0.00
58 T	2-Chloroethyl Vinyl ether	250.000	272.943	-9.2	86	0.00
59 T	2-Hexanone	250.000	241.010	3.6	73	0.00
60 T	Dibromochloromethane	50.000	49.639	0.7	77	0.00
61 T	1,2-Dibromoethane	50.000	52.349	-4.7	81	0.00
62 S	4-Bromofluorobenzene	50.000	51.899	-3.8	82	0.00
63 I	Chlorobenzene-d5	50.000	50.000	0.0	77	0.00
64 T	Tetrachloroethene	50.000	51.075	-2.2	79	0.00
65 PM	Chlorobenzene	50.000	52.036	-4.1	81	0.00
66 T	1,1,1,2-Tetrachloroethane	50.000	51.675	-3.3	80	0.00
67 C	Ethyl Benzene	50.000	51.971	-3.9#	80	0.00
68 T	m/p-Xylenes	100.000	101.212	-1.2	79	0.00
69 T	o-Xylene	50.000	51.506	-3.0	80	0.00
70 T	Styrene	50.000	51.129	-2.3	78	0.00
71 P	Bromoform	50.000	47.958	4.1	73	0.00
72 I	1,4-Dichlorobenzene-d4	50.000	50.000	0.0	76	0.00
73 T	Isopropylbenzene	50.000	51.728	-3.5	80	0.00
74 T	N-ethyl acetate	50.000	50.991	-2.0	78	0.00
75 P	1,1,2,2-Tetrachloroethane	50.000	50.643	-1.3	79	0.00
76 T	1,2,3-Trichloropropane	50.000	52.096	-4.2	79	0.00
77 T	Bromobenzene	50.000	51.674	-3.3	79	0.00
78 T	n-propylbenzene	50.000	52.598	-5.2	79	0.00
79 T	2-Chlorotoluene	50.000	51.128	-2.3	80	0.00
80 T	1,3,5-Trimethylbenzene	50.000	51.457	-2.9	80	0.00
81 T	trans-1,4-Dichloro-2-butene	50.000	47.255	5.5	71	0.00
82 T	4-Chlorotoluene	50.000	50.658	-1.3	79	0.00
83 T	tert-Butylbenzene	50.000	50.656	-1.3	79	0.00
84 T	1,2,4-Trimethylbenzene	50.000	51.952	-3.9	79	0.00
85 T	sec-Butylbenzene	50.000	52.202	-4.4	79	0.00
86 T	p-Isopropyltoluene	50.000	51.345	-2.7	78	0.00
87 T	1,3-Dichlorobenzene	50.000	50.857	-1.7	77	0.00
88 T	1,4-Dichlorobenzene	50.000	50.066	-0.1	78	0.00
89 T	n-Butylbenzene	50.000	50.339	-0.7	77	0.00
90 T	Hexachloroethane	50.000	46.493	7.0	71	0.00
91 T	1,2-Dichlorobenzene	50.000	50.736	-1.5	78	0.00
92 T	1,2-Dibromo-3-Chloropropane	50.000	52.587	-5.2	76	0.00
93 T	1,2,4-Trichlorobenzene	50.000	56.129	-12.3	78	0.00
94 T	Hexachlorobutadiene	50.000	48.998	2.0	76	0.00
95 T	Naphthalene	50.000	55.264	-10.5	80	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112921\
Data File : VX025385.D
Acq On : 29 Nov 2021 21:15
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Nov 30 04:05:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	50.000	57.717	-15.4	81	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6