

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112922\
 Data File : VX033108.D
 Acq On : 29 Nov 2022 11:30
 Operator : JC/MD
 Sample : N5742-04RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-44RE

Quant Time: Nov 30 01:00:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	11405	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	28427	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	14245	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	3492	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	44069	418.529	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 837.060%#			
35) Dibromofluoromethane	5.385	113	26128	163.034	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 326.060%#			
50) Toluene-d8	8.652	98	14698	29.601	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 59.200%#			
62) 4-Bromofluorobenzene	11.079	95	12127	57.605	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 115.220%			
Target Compounds						Qvalue
11) Tert butyl alcohol	2.989	59	23225	1319.123	ug/l	98
16) Acetone	2.391	43	1506	29.034	ug/l	96
19) Methyl tert-butyl Ether	3.117	73	22457	65.061	ug/l	100
20) Methylene Chloride	2.794	84	1811	15.894	ug/l	96
22) Diisopropyl ether	3.763	45	5029	14.373	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112922\
Data File : VX033108.D
Acq On : 29 Nov 2022 11:30
Operator : JC/MD
Sample : N5742-04RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-44RE

Quant Time: Nov 30 01:00:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

