

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112922\
 Data File : VX033116.D
 Acq On : 29 Nov 2022 14:33
 Operator : JC/MD
 Sample : N5742-04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-44

Quant Time: Nov 30 01:03:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	128907	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	205875	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	179366	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	83264	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68080	54.319	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 108.640%	
35) Dibromofluoromethane	5.385	113	54767	47.187	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 94.380%	
50) Toluene-d8	8.647	98	181667	52.112	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.220%	
62) 4-Bromofluorobenzene	11.079	95	65881	43.211	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 86.420%	
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	2.983	59	22217	111.644	ug/l	99
16) Acetone	2.392	43	1269	2.165	ug/l	99
19) Methyl tert-butyl Ether	3.111	73	23630	6.057	ug/l	100
22) Diisopropyl ether	3.764	45	6592	1.667	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112922\
Data File : VX033116.D
Acq On : 29 Nov 2022 14:33
Operator : JC/MD
Sample : N5742-04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-44

Quant Time: Nov 30 01:03:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

