

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112922\
 Data File : VX033118.D
 Acq On : 29 Nov 2022 16:06
 Operator : JC/MD
 Sample : N5805-01 50X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 NWB-1861

Quant Time: Nov 30 01:03:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	143153	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	227512	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	204391	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96991	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	71687	51.332	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.660%
35) Dibromofluoromethane	5.385	113	58617	45.701	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.400%
50) Toluene-d8	8.647	98	195902	50.797	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.600%
62) 4-Bromofluorobenzene	11.079	95	72119	42.804	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.600%
Target Compounds						
					Qvalue	
40) Benzene	6.038	78	1258	0.219	ug/l	95
52) Toluene	8.720	92	3261	0.862	ug/l	98
84) 1,2,4-Trimethylbenzene	11.756	105	3967	0.683	ug/l	94
95) Naphthalene	13.780	128	3644	0.588	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112922\
Data File : VX033118.D
Acq On : 29 Nov 2022 16:06
Operator : JC/MD
Sample : N5805-01 50X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
NWB-1861

Quant Time: Nov 30 01:03:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

