



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX112923\
 Data File : VX039050.D
 Acq On : 29 Nov 2023 09:50
 Operator : JC/MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\624X112423W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Mon Nov 27 03:22:25 2023



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1633

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.1	8372	PASS
75	95	30	60	51.2	22437	PASS
95	95	100	100	100.0	43819	PASS
96	95	5	9	6.8	2964	PASS
173	174	0.00	2	0.7	222	PASS
174	95	50	100	67.9	29757	PASS
175	174	5	9	7.4	2190	PASS
176	174	95	101	97.2	28925	PASS
177	176	5	9	6.5	1885	PASS