

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX113018\
 Data File : VX006221.D
 Acq On : 30 Nov 2018 11:00
 Operator : JC/MD
 Sample : J6134-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRIP-BLANK

Quant Time: Nov 30 22:51:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	1003349	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1467860	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1329092	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	656081	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	469866	45.87	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.74%	
35) Dibromofluoromethane	5.51	113	463162	48.73	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.46%	
50) Toluene-d8	8.71	98	1713854	49.58	ug/l	0.00
Spiked Amount	50.000		Recovery	=	99.16%	
62) 4-Bromofluorobenzene	11.13	95	633871	48.66	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.32%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX113018\
Data File : VX006221.D
Acq On : 30 Nov 2018 11:00
Operator : JC/MD
Sample : J6134-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRIP-BLANK

Quant Time: Nov 30 22:51:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

