

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX113018\
 Data File : VX006227.D
 Acq On : 30 Nov 2018 13:49
 Operator : JC/MD
 Sample : J6147-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-11

Quant Time: Nov 30 23:30:39 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	955255	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1405049	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1274107	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	623280	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	456048	46.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.54%	
35) Dibromofluoromethane	5.51	113	444254	48.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.66%	
50) Toluene-d8	8.71	98	1632707	49.34	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.68%	
62) 4-Bromofluorobenzene	11.13	95	606253	48.62	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.24%	
Target Compounds						
16) Acetone	2.45	43	26797	5.856	ug/l	98
27) cis-1,2-Dichloroethene	4.60	96	26701	2.599	ug/l	90
44) Trichloroethene	7.21	130	14911	1.389	ug/l	83
64) Tetrachloroethene	9.34	164	319523	33.300	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX113018\
Data File : VX006227.D
Acq On : 30 Nov 2018 13:49
Operator : JC/MD
Sample : J6147-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-11

Quant Time: Nov 30 23:30:39 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

