

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX113018\
 Data File : VX006228.D
 Acq On : 30 Nov 2018 14:16
 Operator : JC/MD
 Sample : J6139-01
 Misc : 6.24g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S49-0483

Quant Time: Nov 30 23:37:33 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	836154	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1229674	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1137557	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	619360	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	395158	46.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.58%	
35) Dibromofluoromethane	5.51	113	383684	48.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.38%	
50) Toluene-d8	8.71	98	1425601	49.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.46%	
62) 4-Bromofluorobenzene	11.14	95	560697	51.38	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.76%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.60	96	28340	3.151	ug/l	90
44) Trichloroethene	7.21	130	40339	4.293	ug/l	98
64) Tetrachloroethene	9.34	164	114017	13.309	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX113018\
Data File : VX006228.D
Acq On : 30 Nov 2018 14:16
Operator : JC/MD
Sample : J6139-01
Misc : 6.24g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S49-0483

Quant Time: Nov 30 23:37:33 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

