

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX113018\
 Data File : VX006229.D
 Acq On : 30 Nov 2018 14:43
 Operator : JC/MD
 Sample : J5762-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BU-02-112918

Quant Time: Nov 30 23:41:25 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	919034	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1371813	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1254099	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	632200	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.07	65	446079	47.55	ug/l	0.00
Spiked Amount			Recovery	=	95.10%	
35) Dibromofluoromethane	5.50	113	439684	49.50	ug/l	0.00
Spiked Amount			Recovery	=	99.00%	
50) Toluene-d8	8.71	98	1601755	49.58	ug/l	0.00
Spiked Amount			Recovery	=	99.16%	
62) 4-Bromofluorobenzene	11.13	95	601330	49.39	ug/l	0.00
Spiked Amount			Recovery	=	98.78%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.45	43	52517	11.928	ug/l	97
20) Methylene Chloride	2.86	84	271870	29.014	ug/l	99
25) 2-Butanone	4.70	43	12935	2.059	ug/l	94
43) Isopropyl Acetate	6.46	43	50976	2.363	ug/l	97
64) Tetrachloroethene	9.34	164	82607	8.747	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX113018\
Data File : VX006229.D
Acq On : 30 Nov 2018 14:43
Operator : JC/MD
Sample : J5762-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BU-02-112918

Quant Time: Nov 30 23:41:25 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

