

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX113018\
 Data File : VX006239.D
 Acq On : 30 Nov 2018 19:09
 Operator : JC/MD
 Sample : J6153-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-9.41

Quant Time: Dec 01 00:26:07 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
 Quant Title : SW846 8260
 QLast Update : Fri Nov 30 03:55:33 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	904140	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	1353360	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	1234153	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	606360	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	446452	48.37	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.74%	
35) Dibromofluoromethane	5.51	113	433581	49.48	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.96%	
50) Toluene-d8	8.71	98	1584636	49.72	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.44%	
62) 4-Bromofluorobenzene	11.13	95	587153	48.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.78%	
Target Compounds						
20) Methylene Chloride	2.86	84	9185	0.996	ug/l #	85
64) Tetrachloroethene	9.34	164	15579	1.676	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX113018\
Data File : VX006239.D
Acq On : 30 Nov 2018 19:09
Operator : JC/MD
Sample : J6153-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-9.41

Quant Time: Dec 01 00:26:07 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112918W.M
Quant Title : SW846 8260
QLast Update : Fri Nov 30 03:55:33 2018
Response via : Initial Calibration

