

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX113020\
 Data File : VX019680.D
 Acq On : 30 Nov 2020 15:10
 Operator : JC/MD
 Sample : L4900-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PAV-B2-112520-0-1

Quant Time: Dec 01 03:13:49 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	295057	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	492461	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	420729	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	160053	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	184418	48.21	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.42%		
35) Dibromofluoromethane	5.46	113	148868	48.16	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.32%		
50) Toluene-d8	8.70	98	572946	47.70	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	95.40%		
62) 4-Bromofluorobenzene	11.12	95	189778	43.91	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	87.82%		

Target Compounds

					Qvalue
16) Acetone	2.42	43	13270	8.970	ug/l 99
18) Methyl Acetate	2.76	43	2653	1.357	ug/l 96
20) Methylene Chloride	2.83	84	24126	6.525	ug/l 96
43) Isopropyl Acetate	6.42	43	31195	4.282	ug/l 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX113020\
Data File : VX019680.D
Acq On : 30 Nov 2020 15:10
Operator : JC/MD
Sample : L4900-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAV-B2-112520-0-1

Quant Time: Dec 01 03:13:49 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

