

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX113020\
 Data File : VX019682.D
 Acq On : 30 Nov 2020 15:57
 Operator : JC/MD
 Sample : L4903-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-1-2-COMP

Quant Time: Dec 01 03:17:38 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	281862	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	470077	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	405809	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	161591	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	180228	49.32	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.64%	
35) Dibromofluoromethane		5.46	113	146039	49.49	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.98%	
50) Toluene-d8		8.70	98	561384	48.96	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.92%	
62) 4-Bromofluorobenzene		11.12	95	185555	44.98	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	89.96%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.43	43	14048	9.940	ug/l	99
18) Methyl Acetate	2.75	43	2391	1.316	ug/l	99
20) Methylene Chloride	2.83	84	27588	7.810	ug/l	97
43) Isopropyl Acetate	6.42	43	31636	4.549	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX113020\
Data File : VX019682.D
Acq On : 30 Nov 2020 15:57
Operator : JC/MD
Sample : L4903-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-1-2-COMP

Quant Time: Dec 01 03:17:38 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

