

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX113020\
 Data File : VX019683.D
 Acq On : 30 Nov 2020 16:20
 Operator : JC/MD
 Sample : L4903-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-3-4-COMP

Quant Time: Dec 01 03:19:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	273844	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	461145	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	407346	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	166985	50.00	ug/l	0.00

System Monitoring Compounds		R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4		6.03	65	175520	49.43	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.86%	
35) Dibromofluoromethane		5.46	113	139295	48.12	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.24%	
50) Toluene-d8		8.70	98	546845	48.62	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.24%	
62) 4-Bromofluorobenzene		11.12	95	185541	45.85	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	91.70%	

Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
16) Acetone	2.43	43	15643	11.393	ug/l	96
18) Methyl Acetate	2.75	43	3375	1.624	ug/l #	84
20) Methylene Chloride	2.83	84	26020	7.582	ug/l	98
43) Isopropyl Acetate	6.42	43	29996	4.397	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX113020\
Data File : VX019683.D
Acq On : 30 Nov 2020 16:20
Operator : JC/MD
Sample : L4903-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-3-4-COMP

Quant Time: Dec 01 03:19:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

