

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX113020\
 Data File : VX019684.D
 Acq On : 30 Nov 2020 16:44
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 01 03:20:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	97	0.00
2 T	Dichlorodifluoromethane	0.486	0.465	4.3	90	0.00
3 P	Chloromethane	0.558	0.505	9.5	87	0.00
4 C	Vinyl Chloride	0.601	0.583	3.0#	93	0.00
5 T	Bromomethane	0.227	0.179	21.1	76	0.00
6 T	Chloroethane	0.166	0.134	19.3	82	0.00
7 T	Trichlorofluoromethane	0.867	0.828	4.5	94	0.00
8 T	Diethyl Ether	0.347	0.353	-1.7	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.484	0.489	-1.0	97	0.00
10 T	Methyl Iodide	0.635	0.679	-6.9	99	0.00
11 T	Tert butyl alcohol	0.123	0.150	-22.0	120	0.00
12 CM	1,1-Dichloroethene	0.496	0.492	0.8#	98	0.00
13 T	Acrolein	0.082	0.075	8.5	91	0.00
14 T	Allyl chloride	0.778	0.805	-3.5	104	0.00
15 T	Acrylonitrile	0.289	0.328	-13.5	108	0.00
16 T	Acetone	0.251	0.266	-6.0	105	0.00
17 T	Carbon Disulfide	1.334	1.197	10.3	90	0.00
18 T	Methyl Acetate	0.647	0.712	-10.0	114	0.00
19 T	Methyl tert-butyl Ether	1.713	1.879	-9.7	106	0.00
20 T	Methylene Chloride	0.627	0.608	3.0	99	0.00
21 T	trans-1,2-Dichloroethene	0.562	0.569	-1.2	100	0.00
22 T	Diisopropyl ether	1.609	1.700	-5.7	101	0.00
23 T	Vinyl Acetate	1.349	1.485	-10.1	103	0.00
24 P	1,1-Dichloroethane	0.984	1.038	-5.5	103	0.00
25 T	2-Butanone	0.377	0.422	-11.9	107	0.00
26 T	2,2-Dichloropropane	0.859	0.837	2.6	95	0.00
27 T	cis-1,2-Dichloroethene	0.646	0.679	-5.1	102	0.00
28 T	Bromochloromethane	0.470	0.442	6.0	101	0.00
29 T	Tetrahydrofuran	0.240	0.268	-11.7	108	0.00
30 C	Chloroform	1.029	1.073	-4.3#	100	0.00
31 T	Cyclohexane	0.843	0.832	1.3	96	0.00
32 T	1,1,1-Trichloroethane	0.884	0.924	-4.5	101	0.00
33 S	1,2-Dichloroethane-d4	0.648	0.650	-0.3	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	100	0.00
35 S	Dibromofluoromethane	0.314	0.317	-1.0	103	0.00
36 T	1,1-Dichloropropene	0.469	0.452	3.6	97	0.00
37 T	Ethyl Acetate	0.440	0.483	-9.8	107	0.00
38 T	Carbon Tetrachloride	0.454	0.459	-1.1	99	0.00
39 T	Methylcyclohexane	0.546	0.516	5.5	93	0.00
40 TM	Benzene	1.440	1.427	0.9	98	0.00
41 T	Methacrylonitrile	0.243	0.261	-7.4	105	0.00
42 TM	1,2-Dichloroethane	0.494	0.500	-1.2	100	0.00
43 T	Isopropyl Acetate	0.740	0.781	-5.5	105	0.00
44 TM	Trichloroethene	0.377	0.362	4.0	96	0.00
45 C	1,2-Dichloropropane	0.355	0.363	-2.3#	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX113020\
 Data File : VX019684.D
 Acq On : 30 Nov 2020 16:44
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 01 03:20:41 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
46 T	Dibromomethane	0.243	0.248	-2.1	100	0.00
47 T	Bromodichloromethane	0.471	0.489	-3.8	98	0.00
48 T	Methyl methacrylate	0.346	0.381	-10.1	104	0.00
49 T	1,4-Dioxane	0.008	0.010	-25.0	122	0.00
50 S	Toluene-d8	1.220	1.171	4.0	96	0.00
51 T	4-Methyl-2-Pentanone	0.431	0.489	-13.5	108	0.00
52 CM	Toluene	0.881	0.875	0.7#	95	0.00
53 T	t-1,3-Dichloropropene	0.509	0.531	-4.3	96	0.00
54 T	cis-1,3-Dichloropropene	0.564	0.582	-3.2	97	0.00
55 T	1,1,2-Trichloroethane	0.359	0.365	-1.7	97	0.00
56 T	Ethyl methacrylate	0.508	0.568	-11.8	104	0.00
57 T	1,3-Dichloropropane	0.601	0.619	-3.0	99	0.00
58 T	2-Chloroethyl Vinyl ether	0.298	0.295	1.0	98	0.00
59 T	2-Hexanone	0.327	0.373	-14.1	108	0.00
60 T	Dibromochloromethane	0.351	0.375	-6.8	99	0.00
61 T	1,2-Dibromoethane	0.370	0.381	-3.0	97	0.00
62 S	4-Bromofluorobenzene	0.439	0.448	-2.1	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	97	0.00
64 T	Tetrachloroethene	0.393	0.327	16.8	84	0.00
65 PM	Chlorobenzene	1.028	1.010	1.8	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.374	0.381	-1.9	97	0.00
67 C	Ethyl Benzene	1.801	1.813	-0.7#	94	0.00
68 T	m/p-Xylenes	0.671	0.687	-2.4	95	0.00
69 T	o-Xylene	0.641	0.675	-5.3	97	0.00
70 T	Styrene	1.065	1.149	-7.9	97	0.00
71 P	Bromoform	0.273	0.305	-11.7	104	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	98	0.00
73 T	Isopropylbenzene	3.459	3.524	-1.9	96	0.00
74 T	N-amyl acetate	1.296	1.465	-13.0	107	0.00
75 P	1,1,2,2-Tetrachloroethane	1.237	1.309	-5.8	106	0.00
76 T	1,2,3-Trichloropropane	1.091	1.146	-5.0	103	0.00
77 T	Bromobenzene	0.883	0.898	-1.7	100	0.00
78 T	n-propylbenzene	3.968	3.956	0.3	94	0.00
79 T	2-Chlorotoluene	2.406	2.415	-0.4	96	0.00
80 T	1,3,5-Trimethylbenzene	2.902	3.001	-3.4	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.371	0.408	-10.0	105	0.00
82 T	4-Chlorotoluene	2.833	2.873	-1.4	97	0.00
83 T	tert-Butylbenzene	2.807	2.903	-3.4	98	0.00
84 T	1,2,4-Trimethylbenzene	2.936	3.038	-3.5	95	0.00
85 T	sec-Butylbenzene	3.340	3.379	-1.2	94	0.00
86 T	p-Isopropyltoluene	3.032	3.060	-0.9	93	0.00
87 T	1,3-Dichlorobenzene	1.622	1.619	0.2	97	0.00
88 T	1,4-Dichlorobenzene	1.670	1.639	1.9	98	0.00
89 T	n-Butylbenzene	2.741	2.659	3.0	89	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX113020\
Data File : VX019684.D
Acq On : 30 Nov 2020 16:44
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 01 03:20:41 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
90 T	Hexachloroethane	0.507	0.523	-3.2	97	0.00
91 T	1,2-Dichlorobenzene	1.629	1.613	1.0	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.252	0.287	-13.9	110	0.00
93 T	1,2,4-Trichlorobenzene	1.046	1.048	-0.2	96	0.00
94 T	Hexachlorobutadiene	0.476	0.422	11.3	90	0.00
95 T	Naphthalene	3.237	3.659	-13.0	101	0.00
96 T	1,2,3-Trichlorobenzene	1.046	1.027	1.8	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6