

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\WX113021\
 Data File : VX025398.D
 Acq On : 30 Nov 2021 17:37
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 01 02:22:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	71	0.00
2 T	Dichlorodifluoromethane	0.433	0.445	-2.8	64	0.00
3 P	Chloromethane	0.537	0.517	3.7	66	0.00
4 C	Vinyl Chloride	0.589	0.547	7.1#	62	0.00
5 T	Bromomethane	0.253	0.265	-4.7	73	0.00
6 T	Chloroethane	0.407	0.382	6.1	69	0.00
7 T	Trichlorofluoromethane	0.875	0.812	7.2	64	0.00
8 T	Diethyl Ether	0.361	0.373	-3.3	73	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.465	0.456	1.9	69	0.00
10 T	Methyl Iodide	0.582	0.674	-15.8	75	0.00
11 T	Tert butyl alcohol	0.142	0.122	14.1	64	0.01
12 CM	1,1-Dichloroethene	0.463	0.460	0.6#	70	0.00
13 T	Acrolein	0.024	0.028	-16.7	78	0.00
14 T	Allyl chloride	0.839	0.834	0.6	73	0.00
15 T	Acrylonitrile	0.283	0.303	-7.1	75	0.00
16 T	Acetone	0.293	0.295	-0.7	71	0.00
17 T	Carbon Disulfide	1.304	1.082	17.0	60	0.00
18 T	Methyl Acetate	0.936	0.983	-5.0	77	0.00
19 T	Methyl tert-butyl Ether	1.615	1.784	-10.5	79	0.00
20 T	Methylene Chloride	0.551	0.572	-3.8	79	0.00
21 T	trans-1,2-Dichloroethene	0.508	0.525	-3.3	74	0.00
22 T	Diisopropyl ether	1.597	1.753	-9.8	79	0.00
23 T	Vinyl Acetate	1.283	1.403	-9.4	77	0.00
24 P	1,1-Dichloroethane	0.873	0.949	-8.7	78	0.00
25 T	2-Butanone	0.423	0.440	-4.0	73	0.00
26 T	2,2-Dichloropropane	0.791	0.810	-2.4	72	0.00
27 T	cis-1,2-Dichloroethene	0.593	0.648	-9.3	78	0.00
28 T	Bromochloromethane	0.407	0.453	-11.3	78	0.00
29 T	Tetrahydrofuran	0.262	0.274	-4.6	75	0.00
30 C	Chloroform	0.908	1.025	-12.9#	81	0.00
31 T	Cyclohexane	0.814	0.740	9.1	65	0.00
32 T	1,1,1-Trichloroethane	0.832	0.881	-5.9	76	0.00
33 S	1,2-Dichloroethane-d4	0.527	0.584	-10.8	79	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	73	0.00
35 S	Dibromofluoromethane	0.307	0.323	-5.2	78	0.00
36 T	1,1-Dichloropropene	0.436	0.428	1.8	73	0.00
37 T	Ethyl Acetate	0.496	0.511	-3.0	76	0.00
38 T	Carbon Tetrachloride	0.480	0.474	1.3	72	0.00
39 T	Methylcyclohexane	0.560	0.478	14.6	62	0.00
40 TM	Benzene	1.321	1.361	-3.0	76	0.00
41 T	Methacrylonitrile	0.275	0.283	-2.9	79	0.00
42 TM	1,2-Dichloroethane	0.430	0.482	-12.1	82	0.00
43 T	Isopropyl Acetate	0.826	0.842	-1.9	75	0.00
44 TM	Trichloroethene	0.384	0.388	-1.0	74	0.00
45 C	1,2-Dichloropropane	0.325	0.345	-6.2#	77	0.00
46 T	Dibromomethane	0.236	0.254	-7.6	79	0.00
47 T	Bromodichloromethane	0.463	0.494	-6.7	78	0.00
48 T	Methyl methacrylate	0.388	0.392	-1.0	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113021\
 Data File : VX025398.D
 Acq On : 30 Nov 2021 17:37
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 01 02:22:45 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.008	0.0	66	0.00
50 S	Toluene-d8	1.153	1.164	-1.0	74	0.00
51 T	4-Methyl-2-Pentanone	0.518	0.533	-2.9	75	0.00
52 CM	Toluene	0.863	0.896	-3.8#	76	0.00
53 T	t-1,3-Dichloropropene	0.515	0.543	-5.4	77	0.00
54 T	cis-1,3-Dichloropropene	0.558	0.576	-3.2	75	0.00
55 T	1,1,2-Trichloroethane	0.343	0.381	-11.1	81	0.00
56 T	Ethyl methacrylate	0.556	0.579	-4.1	76	0.00
57 T	1,3-Dichloropropane	0.568	0.617	-8.6	80	0.00
58 T	2-Chloroethyl Vinyl ether	0.259	0.281	-8.5	81	0.00
59 T	2-Hexanone	0.408	0.412	-1.0	72	0.00
60 T	Dibromochloromethane	0.395	0.416	-5.3	77	0.00
61 T	1,2-Dibromoethane	0.371	0.393	-5.9	78	0.00
62 S	4-Bromofluorobenzene	0.436	0.439	-0.7	75	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	74	0.00
64 T	Tetrachloroethene	0.412	0.397	3.6	72	0.00
65 PM	Chlorobenzene	1.019	1.058	-3.8	78	0.00
66 T	1,1,1,2-Tetrachloroethane	0.391	0.409	-4.6	78	0.00
67 C	Ethyl Benzene	1.761	1.800	-2.2#	75	0.00
68 T	m/p-Xylenes	0.719	0.718	0.1	75	0.00
69 T	o-Xylene	0.704	0.714	-1.4	75	0.00
70 T	Styrene	1.163	1.201	-3.3	76	0.00
71 P	Bromoform	0.339	0.330	2.7	71	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	72	0.00
73 T	Isopropylbenzene	3.563	3.582	-0.5	75	0.00
74 T	N-amyl acetate	1.541	1.534	0.5	73	0.00
75 P	1,1,2,2-Tetrachloroethane	1.153	1.189	-3.1	77	0.00
76 T	1,2,3-Trichloropropane	1.024	1.077	-5.2	76	0.00
77 T	Bromobenzene	0.916	0.955	-4.3	77	0.00
78 T	n-propylbenzene	4.000	4.038	-1.0	73	0.00
79 T	2-Chlorotoluene	2.435	2.483	-2.0	77	0.00
80 T	1,3,5-Trimethylbenzene	2.995	3.023	-0.9	75	0.00
81 T	trans-1,4-Dichloro-2-butene	0.406	0.394	3.0	70	0.00
82 T	4-Chlorotoluene	2.780	2.834	-1.9	76	0.00
83 T	tert-Butylbenzene	3.044	2.998	1.5	74	0.00
84 T	1,2,4-Trimethylbenzene	2.960	3.049	-3.0	75	0.00
85 T	sec-Butylbenzene	3.701	3.633	1.8	71	0.00
86 T	p-Isopropyltoluene	3.180	3.142	1.2	72	0.00
87 T	1,3-Dichlorobenzene	1.706	1.742	-2.1	74	0.00
88 T	1,4-Dichlorobenzene	1.729	1.737	-0.5	75	0.00
89 T	n-Butylbenzene	2.680	2.521	5.9	69	0.00
90 T	Hexachloroethane	0.586	0.539	8.0	67	0.00
91 T	1,2-Dichlorobenzene	1.663	1.728	-3.9	76	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.241	0.253	-5.0	73	0.00
93 T	1,2,4-Trichlorobenzene	0.979	1.068	-9.1	73	0.00
94 T	Hexachlorobutadiene	0.470	0.413	12.1	66	0.00
95 T	Naphthalene	3.046	3.609	-18.5	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113021\
Data File : VX025398.D
Acq On : 30 Nov 2021 17:37
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 01 02:22:45 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.957	1.055	-10.2	74	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6