

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113022\
 Data File : VX033126.D
 Acq On : 30 Nov 2022 13:00
 Operator : JC/MD
 Sample : N5801-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TP-13

Quant Time: Dec 01 00:18:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	134552	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	214769	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	205613	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	95405	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	68983	52.632	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	105.260%
35) Dibromofluoromethane	5.385	113	55665	45.974	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.940%
50) Toluene-d8	8.647	98	192298	52.911	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	105.820%
62) 4-Bromofluorobenzene	11.079	95	73280	46.074	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.140%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	19186	31.352	ug/l	99
20) Methylene Chloride	2.788	84	6209	3.674	ug/l	96
37) Ethyl Acetate	4.708	43	14347	8.494	ug/l	98
43) Isopropyl Acetate	6.330	43	68479	24.515	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113022\
Data File : VX033126.D
Acq On : 30 Nov 2022 13:00
Operator : JC/MD
Sample : N5801-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TP-13

Quant Time: Dec 01 00:18:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

