

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113022\
 Data File : VX033127.D
 Acq On : 30 Nov 2022 13:23
 Operator : JC/MD
 Sample : N5803-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PEQ-STOCK

Quant Time: Dec 01 00:19:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	138460	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	227576	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210679	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96655	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	74460	55.371	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.740%
35) Dibromofluoromethane	5.391	113	59690	46.524	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	93.040%
50) Toluene-d8	8.646	98	206268	53.588	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.180%
62) 4-Bromofluorobenzene	11.079	95	76168	45.194	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.380%
Target Compounds						
						Qvalue
16) Acetone	2.391	43	12859	20.420	ug/l	96
20) Methylene Chloride	2.794	84	4715	2.362	ug/l	91
37) Ethyl Acetate	4.720	43	17891	9.996	ug/l	97
43) Isopropyl Acetate	6.336	43	80277	27.121	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113022\
 Data File : VX033127.D
 Acq On : 30 Nov 2022 13:23
 Operator : JC/MD
 Sample : N5803-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PEQ-STOCK

Quant Time: Dec 01 00:19:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

