

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113022\
 Data File : VX033128.D
 Acq On : 30 Nov 2022 13:46
 Operator : JC/MD
 Sample : N5804-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-1-STOCK

Quant Time: Dec 01 00:19:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	145766	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	238191	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	220279	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	97182	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	79173	55.959	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.920%
35) Dibromofluoromethane	5.385	113	63251	47.103	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.200%
50) Toluene-d8	8.647	98	217895	54.107	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	108.220%
62) 4-Bromofluorobenzene	11.079	95	85159	48.277	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.560%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	12187	18.383	ug/l	99
20) Methylene Chloride	2.794	84	4985	2.378	ug/l	96
37) Ethyl Acetate	4.721	43	18846	10.061	ug/l	97
43) Isopropyl Acetate	6.336	43	83637	26.997	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113022\
Data File : VX033128.D
Acq On : 30 Nov 2022 13:46
Operator : JC/MD
Sample : N5804-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-1-STOCK

Quant Time: Dec 01 00:19:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

