

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113022\
 Data File : VX033129.D
 Acq On : 30 Nov 2022 14:09
 Operator : JC/MD
 Sample : N5806-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-1

Quant Time: Dec 01 00:19:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	133870	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	218434	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	206682	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	94045	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	72132	55.486	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.980%
35) Dibromofluoromethane	5.385	113	57915	47.030	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.060%
50) Toluene-d8	8.647	98	198428	53.714	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.420%
62) 4-Bromofluorobenzene	11.079	95	79548	49.175	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.360%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	12690	20.843	ug/l	98
20) Methylene Chloride	2.794	84	4495	2.311	ug/l	96
37) Ethyl Acetate	4.721	43	19307	11.239	ug/l	97
43) Isopropyl Acetate	6.336	43	85432	30.071	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113022\
Data File : VX033129.D
Acq On : 30 Nov 2022 14:09
Operator : JC/MD
Sample : N5806-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-1

Quant Time: Dec 01 00:19:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

