

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113022\
 Data File : VX033133.D
 Acq On : 30 Nov 2022 15:42
 Operator : JC/MD
 Sample : N5809-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HOWE-DRUM-C-1129

Quant Time: Dec 01 00:21:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	148337	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	243121	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	229408	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	102949	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	83646	58.223	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery = 116.440%			
35) Dibromofluoromethane	5.385	113	66506	48.522	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 97.040%			
50) Toluene-d8	8.647	98	230879	56.255	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery = 112.500%#			
62) 4-Bromofluorobenzene	11.079	95	84942	47.178	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery = 94.360%			
Target Compounds					Qvalue	
16) Acetone	2.398	43	15270	22.634	ug/l	93
20) Methylene Chloride	2.788	84	4422	1.902	ug/l	85
37) Ethyl Acetate	4.714	43	20123	10.524	ug/l	99
43) Isopropyl Acetate	6.342	43	90272	28.548	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113022\
Data File : VX033133.D
Acq On : 30 Nov 2022 15:42
Operator : JC/MD
Sample : N5809-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HOWE-DRUM-C-1129

Quant Time: Dec 01 00:21:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

