

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113022\
 Data File : VX033142.D
 Acq On : 30 Nov 2022 19:10
 Operator : JC/MD
 Sample : N5815-11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MW-08

Quant Time: Dec 01 00:24:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	129592	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	211360	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186007	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85541	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	69753	55.424	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.840%
35) Dibromofluoromethane	5.385	113	54542	45.773	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.540%
50) Toluene-d8	8.647	98	182594	50.972	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.940%
62) 4-Bromofluorobenzene	11.079	95	67236	42.955	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.920%
Target Compounds						
						Qvalue
67) Ethyl Benzene	10.195	91	2162	0.329	ug/l #	76
68) m/p-Xylenes	10.305	106	879	0.341	ug/l	96
84) 1,2,4-Trimethylbenzene	11.756	105	2110	0.412	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113022\
Data File : VX033142.D
Acq On : 30 Nov 2022 19:10
Operator : JC/MD
Sample : N5815-11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MW-08

Quant Time: Dec 01 00:24:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

