

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113023\
 Data File : VX039065.D
 Acq On : 30 Nov 2023 14:04
 Operator : JC/MD
 Sample : 05585-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 P3874-SP1

Quant Time: Dec 01 01:57:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	108666	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	222944	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	225064	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	107535	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	84213	48.095	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.180%
35) Dibromofluoromethane	5.385	113	70439	43.747	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	87.500%
50) Toluene-d8	8.647	98	277441	45.035	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	90.080%#
62) 4-Bromofluorobenzene	11.079	95	116954	47.425	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	94.840%
Target Compounds						
16) Acetone	2.374	43	2442	2.620	ug/l	Qvalue # 83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113023\
Data File : VX039065.D
Acq On : 30 Nov 2023 14:04
Operator : JC/MD
Sample : 05585-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
P3874-SP1

Quant Time: Dec 01 01:57:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

