

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113023\
 Data File : VX039069.D
 Acq On : 30 Nov 2023 15:36
 Operator : JC/MD
 Sample : 05585-06
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB

Quant Time: Dec 01 01:59:08 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
 Quant Title : SW846 8260
 QLast Update : Thu Nov 16 22:59:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	112463	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	229100	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	236209	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	119993	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	87373	48.215	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	96.420%
35) Dibromofluoromethane	5.385	113	73909	44.668	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	89.340%
50) Toluene-d8	8.647	98	288625	45.592	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	91.180%#
62) 4-Bromofluorobenzene	11.079	95	124561	49.152	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX113023\
Data File : VX039069.D
Acq On : 30 Nov 2023 15:36
Operator : JC/MD
Sample : 05585-06
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB

Quant Time: Dec 01 01:59:08 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111623W.M
Quant Title : SW846 8260
QLast Update : Thu Nov 16 22:59:22 2023
Response via : Initial Calibration

