

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120120\
 Data File : VX019687.D
 Acq On : 01 Dec 2020 11:50
 Operator : JC/MD
 Sample : VX1201WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampled :
 VX1201WBL01

Quant Time: Dec 02 02:19:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.62	168	306433	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	510962	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	442281	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	175799	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	194706	49.01	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.02%
35) Dibromofluoromethane	5.46	113	154392	48.14	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.28%
50) Toluene-d8	8.70	98	604934	48.54	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.08%
62) 4-Bromofluorobenzene	11.12	95	201780	45.00	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	90.00%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120120\
 Data File : VX019687.D
 Acq On : 01 Dec 2020 11:50
 Operator : JC/MD
 Sample : VX1201WBL01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1201WBL01

Quant Time: Dec 02 02:19:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

