

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120120\
 Data File : VX019690.D
 Acq On : 01 Dec 2020 13:14
 Operator : JC/MD
 Sample : PB133320TB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133320TB

Quant Time: Dec 02 02:25:29 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	283899	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	467396	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	412500	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	169319	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	178323	48.44	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	96.88%		
35) Dibromofluoromethane	5.46	113	143003	48.74	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	97.48%		
50) Toluene-d8	8.70	98	561758	49.27	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.54%		
62) 4-Bromofluorobenzene	11.12	95	191316	46.64	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	93.28%		

Target Compounds

					Qvalue
16) Acetone	2.42	43	10716	7.528	ug/l 98
18) Methyl Acetate	2.75	43	1672	1.107	ug/l 96
20) Methylene Chloride	2.83	84	22278	6.262	ug/l 97
43) Isopropyl Acetate	6.42	43	32956	4.766	ug/l 100
95) Naphthalene	13.82	128	2600	2.377	ug/l 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120120\
Data File : VX019690.D
Acq On : 01 Dec 2020 13:14
Operator : JC/MD
Sample : PB133320TB
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133320TB

Quant Time: Dec 02 02:25:29 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

