

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120120\
 Data File : VX019694.D
 Acq On : 01 Dec 2020 14:47
 Operator : JC/MD
 Sample : PB133320ZHE#07
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB133320ZHE#07

Quant Time: Dec 02 03:04:13 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	286106	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	478552	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	422018	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	155839	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	181054	48.81	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	97.62%
35) Dibromofluoromethane	5.46	113	144701	48.17	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.34%
50) Toluene-d8	8.70	98	576332	49.37	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	98.74%
62) 4-Bromofluorobenzene	11.12	95	185959	44.28	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	88.56%

Target Compounds

					Qvalue
16) Acetone	2.43	43	10837	7.554	ug/l 96
18) Methyl Acetate	2.75	43	3106	1.506	ug/l 93
20) Methylene Chloride	2.83	84	23333	6.507	ug/l 97
43) Isopropyl Acetate	6.42	43	34107	4.817	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120120\
Data File : VX019694.D
Acq On : 01 Dec 2020 14:47
Operator : JC/MD
Sample : PB133320ZHE#07
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB133320ZHE#07

Quant Time: Dec 02 03:04:13 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

