

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120120\
 Data File : VX019697.D
 Acq On : 01 Dec 2020 15:58
 Operator : JC/MD
 Sample : L4910-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PAV-B47-113020-0-1

Quant Time: Dec 02 03:20:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	283351	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	471520	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	407393	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	169684	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	179135	48.76	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.52%		
35) Dibromofluoromethane	5.46	113	143283	48.41	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.82%		
50) Toluene-d8	8.70	98	557594	48.48	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	96.96%		
62) 4-Bromofluorobenzene	11.12	95	189962	45.91	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.82%		

Target Compounds

					Qvalue
16) Acetone	2.42	43	18508	13.027	ug/l 99
18) Methyl Acetate	2.75	43	2396	1.314	ug/l 92
20) Methylene Chloride	2.83	84	22208	6.254	ug/l 93
25) 2-Butanone	4.65	43	14319	6.709	ug/l 96
43) Isopropyl Acetate	6.42	43	31749	4.551	ug/l 100
95) Naphthalene	13.82	128	58619	6.658	ug/l 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120120\
Data File : VX019697.D
Acq On : 01 Dec 2020 15:58
Operator : JC/MD
Sample : L4910-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAV-B47-113020-0-1

Quant Time: Dec 02 03:20:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

