

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX120120\
 Data File : VX019698.D
 Acq On : 01 Dec 2020 16:21
 Operator : JC/MD
 Sample : L4910-08
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PAV-B47-113020-12-12.75

Quant Time: Dec 02 03:21:53 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
 Quant Title : SW846 8260
 QLast Update : Sat Nov 21 01:57:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	295567	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.83	114	497208	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	438577	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	174386	50.00	ug/l	0.00

System Monitoring Compounds

33) 1,2-Dichloroethane-d4	6.03	65	187717	48.98	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	97.96%		
35) Dibromofluoromethane	5.46	113	150147	48.11	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	96.22%		
50) Toluene-d8	8.70	98	594489	49.02	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	98.04%		
62) 4-Bromofluorobenzene	11.12	95	198944	45.59	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	91.18%		

Target Compounds

					Qvalue
16) Acetone	2.42	43	17315	11.683	ug/l 98
18) Methyl Acetate	2.76	43	2441	1.298	ug/l 97
20) Methylene Chloride	2.83	84	20960	5.659	ug/l 98
25) 2-Butanone	4.64	43	13447	6.040	ug/l 97
43) Isopropyl Acetate	6.41	43	31552	4.289	ug/l 99
95) Naphthalene	13.82	128	11670	3.046	ug/l 98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX120120\
Data File : VX019698.D
Acq On : 01 Dec 2020 16:21
Operator : JC/MD
Sample : L4910-08
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PAV-B47-113020-12-12.75

Quant Time: Dec 02 03:21:53 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X112120W.M
Quant Title : SW846 8260
QLast Update : Sat Nov 21 01:57:45 2020
Response via : Initial Calibration

