

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120121\
 Data File : VX025402.D
 Acq On : 01 Dec 2021 10:40
 Operator : JC/MD
 Sample : VX1201WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1201WBL01

Quant Time: Dec 02 08:25:47 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	125870	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	206827	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	183326	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	82315	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	70193	52.906	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 105.820%	
35) Dibromofluoromethane	5.385	113	63765	50.247	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.500%	
50) Toluene-d8	8.647	98	237357	49.746	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.500%	
62) 4-Bromofluorobenzene	11.079	95	81544	45.175	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 90.360%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120121\
Data File : VX025402.D
Acq On : 01 Dec 2021 10:40
Operator : JC/MD
Sample : VX1201WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1201WBL01

Quant Time: Dec 02 08:25:47 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

