



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120121\
Data File : VX025399.D
Acq On : 01 Dec 2021 08:58
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Title : SW846 8260
Last Update : Wed Nov 17 15:36:16 2021



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1632

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.5	4786	PASS
75	95	30	60	49.1	12736	PASS
95	95	100	100	100.0	25920	PASS
96	95	5	9	7.0	1806	PASS
173	174	0.00	2	0.6	129	PASS
174	95	50	100	81.2	21050	PASS
175	174	5	9	7.6	1605	PASS
176	174	95	101	96.6	20324	PASS
177	176	5	9	6.4	1301	PASS