

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
 Data File : VX033157.D
 Acq On : 01 Dec 2022 12:50
 Operator : JC/MD
 Sample : N5815-27
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-2022-11-29

Quant Time: Dec 02 00:50:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	154537	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	246593	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	210731	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	96286	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	75697	50.137	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.280%
35) Dibromofluoromethane	5.385	113	63314	45.543	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.080%
50) Toluene-d8	8.647	98	212383	50.810	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.620%
62) 4-Bromofluorobenzene	11.079	95	78337	42.897	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	85.800%
Target Compounds						
						Qvalue
16) Acetone	2.398	43	1049	1.493	ug/l #	88
67) Ethyl Benzene	10.195	91	10342	1.388	ug/l	98
68) m/p-Xylenes	10.305	106	4083	1.399	ug/l	97
78) n-propylbenzene	11.305	91	5275	0.671	ug/l	97
80) 1,3,5-Trimethylbenzene	11.451	105	1334	0.231	ug/l	94
84) 1,2,4-Trimethylbenzene	11.756	105	9496	1.648	ug/l	98
89) n-Butylbenzene	12.329	91	1653	0.329	ug/l #	61
95) Naphthalene	13.780	128	6833	1.110	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
Data File : VX033157.D
Acq On : 01 Dec 2022 12:50
Operator : JC/MD
Sample : N5815-27
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-2022-11-29

Quant Time: Dec 02 00:50:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

