

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
 Data File : VX033158.D
 Acq On : 01 Dec 2022 13:13
 Operator : JC/MD
 Sample : N5815-28
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-2022-11-29

Quant Time: Dec 02 00:50:42 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	139185	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	220899	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	191114	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	88108	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	72987	53.910	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.820%	
35) Dibromofluoromethane	5.391	113	57815	46.425	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 92.840%	
50) Toluene-d8	8.647	98	192190	51.350	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.700%	
62) 4-Bromofluorobenzene	11.079	95	70274	42.958	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 85.920%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
Data File : VX033158.D
Acq On : 01 Dec 2022 13:13
Operator : JC/MD
Sample : N5815-28
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-2022-11-29

Quant Time: Dec 02 00:50:42 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

