

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
 Data File : VX033159.D
 Acq On : 01 Dec 2022 13:36
 Operator : JC/MD
 Sample : N5828-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-02-11302022

Quant Time: Dec 02 00:50:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	133285	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	211714	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	186413	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	86907	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71027	54.839	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.680%	
35) Dibromofluoromethane	5.385	113	57679	48.325	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 96.640%	
50) Toluene-d8	8.647	98	188104	52.486	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.980%	
62) 4-Bromofluorobenzene	11.079	95	70292	44.833	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 89.660%	
Target Compounds						Qvalue
16) Acetone	2.392	43	1775	2.928	ug/l	# 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
Data File : VX033159.D
Acq On : 01 Dec 2022 13:36
Operator : JC/MD
Sample : N5828-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-02-11302022

Quant Time: Dec 02 00:50:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

