

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
 Data File : VX033162.D
 Acq On : 01 Dec 2022 14:46
 Operator : JC/MD
 Sample : N5828-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 RW9-PT-IN-02-11302022

Quant Time: Dec 02 00:52:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126226	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	201988	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	175930	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	81259	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68268	55.706	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 111.420%	
35) Dibromofluoromethane	5.385	113	54378	47.753	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.500%	
50) Toluene-d8	8.647	98	178447	52.176	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 104.360%	
62) 4-Bromofluorobenzene	11.079	95	65575	43.838	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 87.680%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
Data File : VX033162.D
Acq On : 01 Dec 2022 14:46
Operator : JC/MD
Sample : N5828-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
RW9-PT-IN-02-11302022

Quant Time: Dec 02 00:52:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

