

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
 Data File : VX033164.D
 Acq On : 01 Dec 2022 15:32
 Operator : JC/MD
 Sample : N5815-27
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EB-2022-11-29

Quant Time: Dec 02 00:52:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	122099	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	198451	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	169955	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	78765	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68257	57.692	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	115.380%
35) Dibromofluoromethane	5.385	113	54524	48.735	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.460%
50) Toluene-d8	8.647	98	178221	53.076	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	106.160%
62) 4-Bromofluorobenzene	11.079	95	66027	44.927	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.860%
Target Compounds						
16) Acetone	2.404	43	659	1.187	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
Data File : VX033164.D
Acq On : 01 Dec 2022 15:32
Operator : JC/MD
Sample : N5815-27
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EB-2022-11-29

Quant Time: Dec 02 00:52:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

