

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
 Data File : VX033165.D
 Acq On : 01 Dec 2022 15:55
 Operator : JC/MD
 Sample : N5821-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22-MW-3I

Quant Time: Dec 02 00:52:59 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	119859	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	190736	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	166274	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76468	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	65721	56.523	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	113.040%
35) Dibromofluoromethane	5.385	113	53048	49.333	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.660%
50) Toluene-d8	8.647	98	168922	52.311	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	104.620%
62) 4-Bromofluorobenzene	11.079	95	62367	44.153	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	88.300%
Target Compounds						
						Qvalue
27) cis-1,2-Dichloroethene	4.501	96	700	0.515	ug/l	79
44) Trichloroethene	7.122	130	3677	2.718	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
Data File : VX033165.D
Acq On : 01 Dec 2022 15:55
Operator : JC/MD
Sample : N5821-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22-MW-3I

Quant Time: Dec 02 00:52:59 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

