

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
 Data File : VX033166.D
 Acq On : 01 Dec 2022 16:18
 Operator : JC/MD
 Sample : N5821-03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22-MW-3S

Quant Time: Dec 02 00:53:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	120573	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	192205	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	170669	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	79629	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	64573	55.129	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 110.260%	
35) Dibromofluoromethane	5.385	113	51770	47.777	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 95.560%	
50) Toluene-d8	8.647	98	168749	51.838	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 103.680%	
62) 4-Bromofluorobenzene	11.079	95	62593	43.974	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 87.940%	
Target Compounds						
						Qvalue
16) Acetone	2.385	43	778	1.419	ug/l	98
27) cis-1,2-Dichloroethene	4.495	96	15862	11.599	ug/l	90
32) 1,1,1-Trichloroethane	5.385	97	1584	0.746	ug/l #	48
44) Trichloroethene	7.129	130	996447	730.993	ug/l	97
64) Tetrachloroethene	9.274	164	279	0.231	ug/l #	65

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
Data File : VX033166.D
Acq On : 01 Dec 2022 16:18
Operator : JC/MD
Sample : N5821-03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22-MW-3S

Quant Time: Dec 02 00:53:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

