

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
 Data File : VX033168.D
 Acq On : 01 Dec 2022 17:04
 Operator : JC/MD
 Sample : N5821-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22-MW-6S

Quant Time: Dec 02 00:53:49 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	125911	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	201318	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	176988	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	80265	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	66881	54.651	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.300%	
35) Dibromofluoromethane	5.385	113	52542	46.294	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 92.580%	
50) Toluene-d8	8.647	98	171962	50.373	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 100.740%	
62) 4-Bromofluorobenzene	11.079	95	63250	42.425	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 84.840%	
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.374	62	610240	527.756	ug/l	100
12) 1,1-Dichloroethene	2.325	96	383	0.349	ug/l #	74
21) trans-1,2-Dichloroethene	3.105	96	446	0.355	ug/l #	73
27) cis-1,2-Dichloroethene	4.489	96	427094	299.073	ug/l	91
44) Trichloroethene	7.129	130	11157	7.814	ug/l	96
52) Toluene	8.726	92	1583	0.473	ug/l	97
69) o-Xylene	10.640	106	1435	0.605	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
Data File : VX033168.D
Acq On : 01 Dec 2022 17:04
Operator : JC/MD
Sample : N5821-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22-MW-6S

Quant Time: Dec 02 00:53:49 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

