

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
 Data File : VX033169.D
 Acq On : 01 Dec 2022 17:27
 Operator : JC/MD
 Sample : N5821-09
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 LEA-8I

Quant Time: Dec 02 00:54:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	126695	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	204609	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	181610	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	85389	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	68061	55.309	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	110.620%
35) Dibromofluoromethane	5.385	113	54651	47.378	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	94.760%
50) Toluene-d8	8.647	98	178501	51.495	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	103.000%
62) 4-Bromofluorobenzene	11.079	95	67851	44.779	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.560%
Target Compounds						
						Qvalue
4) Vinyl Chloride	1.374	62	1377	1.184	ug/l #	70
12) 1,1-Dichloroethene	2.319	96	843	0.762	ug/l	99
21) trans-1,2-Dichloroethene	3.099	96	1483	1.173	ug/l	93
27) cis-1,2-Dichloroethene	4.489	96	225261	156.763	ug/l	91
30) Chloroform	5.099	83	1600	0.642	ug/l	96
32) 1,1,1-Trichloroethane	5.391	97	18000	8.069	ug/l	98
38) Carbon Tetrachloride	5.684	117	909	0.444	ug/l #	76
44) Trichloroethene	7.153	130	12497301	8612.220	ug/l	97
64) Tetrachloroethene	9.275	164	4139	3.220	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
Data File : VX033169.D
Acq On : 01 Dec 2022 17:27
Operator : JC/MD
Sample : N5821-09
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
LEA-8I

Quant Time: Dec 02 00:54:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

