

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
 Data File : VX033170.D
 Acq On : 01 Dec 2022 17:50
 Operator : JC/MD
 Sample : N5807-05RE
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 COMP-3RE

Quant Time: Dec 02 00:54:23 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	129202	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	212558	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	199055	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	90945	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	71477	57.058	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	114.120%
35) Dibromofluoromethane	5.385	113	57220	47.750	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	95.500%
50) Toluene-d8	8.647	98	192786	53.626	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	107.260%
62) 4-Bromofluorobenzene	11.079	95	72760	46.223	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.440%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	12624	21.483	ug/l	96
20) Methylene Chloride	2.788	84	4025	2.047	ug/l #	90
37) Ethyl Acetate	4.715	43	18998	11.365	ug/l	99
43) Isopropyl Acetate	6.342	43	85263	30.841	ug/l	100
44) Trichloroethene	7.129	130	22835	15.148	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
Data File : VX033170.D
Acq On : 01 Dec 2022 17:50
Operator : JC/MD
Sample : N5807-05RE
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
COMP-3RE

Quant Time: Dec 02 00:54:23 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

