

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
 Data File : VX033147.D
 Acq On : 01 Dec 2022 08:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 02 00:42:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	106	0.00
2 T	Dichlorodifluoromethane	0.382	0.326	14.7	94	0.00
3 P	Chloromethane	0.412	0.339	17.7	92	0.00
4 C	Vinyl Chloride	0.459	0.406	11.5#	96	0.00
5 T	Bromomethane	0.340	0.315	7.4	105	0.01
6 T	Chloroethane	0.374	0.356	4.8	100	0.00
7 T	Trichlorofluoromethane	0.891	0.900	-1.0	110	0.00
8 T	Diethyl Ether	0.281	0.289	-2.8	111	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.468	0.471	-0.6	107	0.00
10 T	Methyl Iodide	0.536	0.398	25.7#	73	0.00
11 T	Tert butyl alcohol	0.077	0.077	0.0	106	0.04
12 CM	1,1-Dichloroethene	0.436	0.419	3.9#	102	0.00
13 T	Acrolein	0.089	0.092	-3.4	115	0.00
14 T	Allyl chloride	0.704	0.686	2.6	102	0.00
15 T	Acrylonitrile	0.218	0.208	4.6	103	0.00
16 T	Acetone	0.227	0.255	-12.3	119	0.01
17 T	Carbon Disulfide	1.205	1.002	16.8	92	0.00
18 T	Methyl Acetate	0.629	0.610	3.0	106	0.00
19 T	Methyl tert-butyl Ether	1.513	1.544	-2.0	107	0.00
20 T	Methylene Chloride	0.548	0.476	13.1	103	0.00
21 T	trans-1,2-Dichloroethene	0.499	0.465	6.8	101	0.00
22 T	Diisopropyl ether	1.534	1.622	-5.7	107	0.00
23 T	Vinyl Acetate	1.201	1.243	-3.5	104	0.00
24 P	1,1-Dichloroethane	0.878	0.868	1.1	105	0.00
25 T	2-Butanone	0.314	0.332	-5.7	111	0.00
26 T	2,2-Dichloropropane	0.706	0.766	-8.5	112	0.00
27 T	cis-1,2-Dichloroethene	0.567	0.576	-1.6	105	0.00
28 T	Bromochloromethane	0.386	0.381	1.3	102	0.00
29 T	Tetrahydrofuran	0.199	0.192	3.5	104	0.00
30 C	Chloroform	0.983	1.024	-4.2#	108	0.00
31 T	Cyclohexane	0.789	0.768	2.7	101	0.00
32 T	1,1,1-Trichloroethane	0.880	0.930	-5.7	108	0.00
33 S	1,2-Dichloroethane-d4	0.539	0.454	15.8	98	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	106	0.00
35 S	Dibromofluoromethane	0.282	0.256	9.2	101	0.00
36 T	1,1-Dichloropropene	0.451	0.461	-2.2	106	0.00
37 T	Ethyl Acetate	0.393	0.393	0.0	103	0.00
38 T	Carbon Tetrachloride	0.500	0.522	-4.4	108	0.00
39 T	Methylcyclohexane	0.514	0.527	-2.5	105	0.00
40 TM	Benzene	1.264	1.290	-2.1	105	0.00
41 T	Methacrylonitrile	0.221	0.226	-2.3	105	0.00
42 TM	1,2-Dichloroethane	0.509	0.530	-4.1	106	0.00
43 T	Isopropyl Acetate	0.650	0.670	-3.1	105	0.00
44 TM	Trichloroethene	0.355	0.363	-2.3	104	0.00
45 C	1,2-Dichloropropane	0.312	0.334	-7.1#	108	0.00
46 T	Dibromomethane	0.244	0.248	-1.6	105	0.00
47 T	Bromodichloromethane	0.471	0.520	-10.4	111	0.00
48 T	Methyl methacrylate	0.335	0.347	-3.6	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
 Data File : VX033147.D
 Acq On : 01 Dec 2022 08:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Dec 02 00:42:24 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
 Quant Title : SW846 8260
 QLast Update : Tue Nov 22 13:58:46 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.006	0.006	0.0	108	0.04
50 S	Toluene-d8	0.915	0.766	16.3	97	0.00
51 T	4-Methyl-2-Pentanone	0.401	0.416	-3.7	105	0.00
52 CM	Toluene	0.831	0.869	-4.6#	106	0.00
53 T	t-1,3-Dichloropropene	0.462	0.528	-14.3	108	0.00
54 T	cis-1,3-Dichloropropene	0.508	0.561	-10.4	109	0.00
55 T	1,1,2-Trichloroethane	0.338	0.353	-4.4	108	0.00
56 T	Ethyl methacrylate	0.466	0.511	-9.7	106	0.00
57 T	1,3-Dichloropropane	0.546	0.572	-4.8	107	0.00
58 T	2-Chloroethyl Vinyl ether	0.221	0.229	-3.6	105	0.00
59 T	2-Hexanone	0.296	0.325	-9.8	108	0.00
60 T	Dibromochloromethane	0.364	0.402	-10.4	107	0.00
61 T	1,2-Dibromoethane	0.357	0.378	-5.9	106	0.00
62 S	4-Bromofluorobenzene	0.370	0.353	4.6	102	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	106	0.00
64 T	Tetrachloroethene	0.354	0.338	4.5	102	0.00
65 PM	Chlorobenzene	0.997	1.027	-3.0	108	0.00
66 T	1,1,1,2-Tetrachloroethane	0.365	0.399	-9.3	110	0.00
67 C	Ethyl Benzene	1.768	1.873	-5.9#	106	0.00
68 T	m/p-Xylenes	0.692	0.732	-5.8	107	0.00
69 T	o-Xylene	0.670	0.715	-6.7	107	0.00
70 T	Styrene	1.096	1.198	-9.3	106	0.00
71 P	Bromoform	0.280	0.312	-11.4	109	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.573	3.658	-2.4	108	0.00
74 T	N-amyl acetate	1.215	1.256	-3.4	104	0.00
75 P	1,1,2,2-Tetrachloroethane	1.078	1.062	1.5	107	0.00
76 T	1,2,3-Trichloropropane	0.947	0.987	-4.2	112	0.00
77 T	Bromobenzene	0.879	0.856	2.6	105	0.00
78 T	n-propylbenzene	4.085	4.248	-4.0	108	0.00
79 T	2-Chlorotoluene	2.481	2.506	-1.0	108	0.00
80 T	1,3,5-Trimethylbenzene	3.002	3.124	-4.1	107	0.00
81 T	trans-1,4-Dichloro-2-butene	0.291	0.309	-6.2	109	0.00
82 T	4-Chlorotoluene	2.908	2.952	-1.5	107	0.00
83 T	tert-Butylbenzene	2.988	3.068	-2.7	107	0.00
84 T	1,2,4-Trimethylbenzene	2.993	3.092	-3.3	106	0.00
85 T	sec-Butylbenzene	3.574	3.762	-5.3	108	0.00
86 T	p-Isopropyltoluene	3.028	3.203	-5.8	108	0.00
87 T	1,3-Dichlorobenzene	1.648	1.673	-1.5	109	0.00
88 T	1,4-Dichlorobenzene	1.688	1.697	-0.5	107	0.00
89 T	n-Butylbenzene	2.609	2.788	-6.9	109	0.00
90 T	Hexachloroethane	0.498	0.525	-5.4	110	0.00
91 T	1,2-Dichlorobenzene	1.603	1.623	-1.2	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.215	0.226	-5.1	107	0.00
93 T	1,2,4-Trichlorobenzene	0.988	1.037	-5.0	111	0.00
94 T	Hexachlorobutadiene	0.436	0.429	1.6	110	0.00
95 T	Naphthalene	3.197	3.291	-2.9	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120122\
Data File : VX033147.D
Acq On : 01 Dec 2022 08:48
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Dec 02 00:42:24 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X112222W.M
Quant Title : SW846 8260
QLast Update : Tue Nov 22 13:58:46 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.991	1.019	-2.8	108	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6