

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
 Data File : VX025418.D
 Acq On : 02 Dec 2021 10:57
 Operator : JC/MD
 Sample : VX1202WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX1202WBL01

Quant Time: Dec 03 03:03:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	117913	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	197045	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	175310	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	77517	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	67305	54.152	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.300%
35) Dibromofluoromethane	5.391	113	60293	49.870	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.740%
50) Toluene-d8	8.653	98	225241	49.550	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.100%
62) 4-Bromofluorobenzene	11.079	95	77546	45.093	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	90.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
Data File : VX025418.D
Acq On : 02 Dec 2021 10:57
Operator : JC/MD
Sample : VX1202WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX1202WBL01

Quant Time: Dec 03 03:03:58 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

