

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
 Data File : VX025421.D
 Acq On : 02 Dec 2021 12:22
 Operator : JC/MD
 Sample : M4908-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TW-1

Quant Time: Dec 03 03:05:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	114100	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	190687	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	170743	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73155	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	65879	54.776	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	109.560%
35) Dibromofluoromethane	5.391	113	58004	49.576	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.160%
50) Toluene-d8	8.653	98	219141	49.816	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.640%
62) 4-Bromofluorobenzene	11.085	95	74845	44.974	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	89.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
Data File : VX025421.D
Acq On : 02 Dec 2021 12:22
Operator : JC/MD
Sample : M4908-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TW-1

Quant Time: Dec 03 03:05:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

