

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
 Data File : VX025423.D
 Acq On : 02 Dec 2021 13:09
 Operator : JC/MD
 Sample : M4853-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TB-01-112921

Quant Time: Dec 03 03:06:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	112429	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	189655	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	169069	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	75930	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	65830	55.549	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	111.100%
35) Dibromofluoromethane	5.391	113	58808	50.537	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.080%
50) Toluene-d8	8.653	98	217559	49.725	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	99.460%
62) 4-Bromofluorobenzene	11.085	95	76662	46.316	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	92.640%
Target Compounds						
16) Acetone	2.380	43	761	1.156	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
Data File : VX025423.D
Acq On : 02 Dec 2021 13:09
Operator : JC/MD
Sample : M4853-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TB-01-112921

Quant Time: Dec 03 03:06:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

