

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
 Data File : VX025424.D
 Acq On : 02 Dec 2021 13:33
 Operator : JC/MD
 Sample : M4853-01DL 10X
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GWBR3D-BA-472-493-112921DL

Quant Time: Dec 03 03:06:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	112072	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	187714	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	167808	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76142	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	64632	54.712	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 109.420%	
35) Dibromofluoromethane	5.391	113	57928	50.296	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.600%	
50) Toluene-d8	8.653	98	215933	49.864	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 99.720%	
62) 4-Bromofluorobenzene	11.085	95	75477	46.072	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 92.140%	
Target Compounds						
					Qvalue	
12) 1,1-Dichloroethene	2.318	96	1080	1.040	ug/l	86
16) Acetone	2.379	43	1839	2.802	ug/l	90
27) cis-1,2-Dichloroethene	4.495	96	70599	53.128	ug/l	80
44) Trichloroethene	7.129	130	13151	9.125	ug/l	96
52) Toluene	8.720	92	4286	1.323	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
Data File : VX025424.D
Acq On : 02 Dec 2021 13:33
Operator : JC/MD
Sample : M4853-01DL 10X
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GWBR3D-BA-472-493-112921DL

Quant Time: Dec 03 03:06:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

