

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
 Data File : VX025426.D
 Acq On : 02 Dec 2021 14:21
 Operator : JC/MD
 Sample : PB141130ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141130ZHE#01

Quant Time: Dec 03 03:06:56 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	113478	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	190875	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	171570	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	76081	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	66119	55.277	ug/l	0.00
Spiked Amount	50.000					

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
Data File : VX025426.D
Acq On : 02 Dec 2021 14:21
Operator : JC/MD
Sample : PB141130ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB141130ZHE#01

Quant Time: Dec 03 03:06:56 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
Quant Title : SW846 8260
QLast Update : Wed Nov 17 15:36:16 2021
Response via : Initial Calibration

