

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
 Data File : VX025427.D
 Acq On : 02 Dec 2021 14:45
 Operator : JC/MD
 Sample : PB141130ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141130ZHE#02

Quant Time: Dec 03 03:07:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	110854	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.769	114	183144	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	164211	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	73080	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	63482	54.329	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	108.660%
35) Dibromofluoromethane	5.391	113	56630	50.396	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.800%
50) Toluene-d8	8.653	98	211833	50.138	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.280%
62) 4-Bromofluorobenzene	11.085	95	73371	45.904	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	91.800%
Target Compounds						
20) Methylene Chloride	2.794	84	3978	3.259	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX120221\
 Data File : VX025427.D
 Acq On : 02 Dec 2021 14:45
 Operator : JC/MD
 Sample : PB141130ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB141130ZHE#02

Quant Time: Dec 03 03:07:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X111721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 17 15:36:16 2021
 Response via : Initial Calibration

